

Ilgner et al. *Head & Neck Oncology* 2010, **2**(Suppl 1):O30  
<http://www.headandneckoncology.org/content/2/S1/O30>



Head & Neck  
Oncology

ORAL PRESENTATION

Open Access

# Femtosecond laser microstructuring and bioactive nanocoating of titanium surfaces in relation to chondrocyte growth

Justus F Ilgner, Slavomir Biedron, Elena Fadeeva\*, Boris Cichkov, Doris Klee, Anneke Loos, Eveline Sowa-Soehle, Martin Westhofen

From 2nd Scientific Meeting of the Head and Neck Optical Diagnostics Society  
San Francisco, CA, USA. 23-24 January 2010

## Introduction

Titanium implants can be regarded as the current gold standard for restoration of sound transmission in the middle ear following destruction of the ossicular chain by chronic inflammation. Many efforts have been made to improve prosthesis design, while less attention had been given to the role of the interface. We present a study on chemical nanocoating on microstructured titanium contact surface with bioactive protein.

## Materials and methods

Titanium samples of 5mm diameter and 0,25mm thickness were structured by means of a Ti:Sapphire femtosecond laser operating at 970nm with parallel lines of 5µm depth, 5µm width and 10µm inter-groove distance. In addition, various nanolayers were applied to PVDF foils, while bone matrix protein 7 (BMP-7) was linked to the outer coating layer.

## Results

Chondrocytes could be cultured on microstructured surfaces without reduced rate of vital / dead cells compared to native surfaces. Chondrocytes also showed contact guidance by growing along ridges particularly on 5µm lines. On PVDF foils, chondrocyte growth was doubled in contact with BMP-7 compared to hydrogel layer or native surface.

## Discussion

According to these results, relative preference for cell growth on titanium prosthesis contact surfaces compared to non-contact surfaces (e.g. prosthesis shaft) can

be achieved by nanocoating. Relative selectivity induced by microstructures for growth of chondrocytes compared to fibrocytes is subject to further evaluation.

Published: 29 October 2010

doi:10.1186/1758-3284-2-S1-O30

**Cite this article as:** Ilgner et al.: Femtosecond laser microstructuring and bioactive nanocoating of titanium surfaces in relation to chondrocyte growth. *Head & Neck Oncology* 2010 **2**(Suppl 1):O30.

**Submit your next manuscript to BioMed Central and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at  
[www.biomedcentral.com/submit](http://www.biomedcentral.com/submit)



Laser Zentrum Hannover e.V., Hannover, Germany



© 2010 Fadeeva et al; licensee BioMed Central Ltd.